

**The AAMI Standard 6 Pin patient ECG inputs referred to in our
cable descriptions connect to the devices listed below:**

Manufacturer	Type	Model
Critikon	Monitors	Dynamap MPS, 8700 & 9700 series
Datascope Corp.	Balloon Pumps	All from System 90 and later
Datascope Corp.	Defibrillators	All
Datascope Corp.	Monitors	All except Expert series
Invivo Research	Monitors	Millennia series, Precess, Magnitude
Marquette	Defibrillators	900, 1200, 1500
MDE (now Invivo)	Monitors	All
Philips/HP	Defibrillators	Customer Option Only
Philips/HP	Monitors	Customer Option Only
Physio-Control	Defibrillators	LP5, LP6, LP7, LP9, LP10
PLC Medical Systems	CO2 Heart Laser	All
Quinton	Hemodynamic Monitor	Q-Cath
Spacelabs	Monitors	90300 series, 90400 series, 90600 series, 90700 series except Ultraview series 90367, 90369, 90496
WelchAllyn	Monitors	Atlas
WelchAllyn/MRL	Defibrillators	450SL, 360SLx, PIC
WelchAllyn/Protocol	Monitors	All
Zoll Medical	Defibrillators	All

MaguireEnterprises.com

1-800-548-9686